


PainShield Application for Trigeminal Neuralgia


PainShield use should start gradually with the first and second treatments lasting 2 and 4 hours respectively during day time. Following treatments should be for 6.5 hours on a daily basis and may be done overnight and the PainShield removed upon waking. The device is programmed to work in cycles of 30 minutes on and 30 minutes off, for a total of 6.5 hours of intermittent treatment.




Connect the actuator cable to the device socket.



Remove the protective liner from the center portion of the patch only.



Attach the actuator to the exposed adhesive, the wire aligned with the open notch



Remove the rest of the protective liner from the patch.

Place and adhere the facial patch so that it is abutted next to a bony prominence in the painful region; just below the cheek bone, adjacent to the ear. In cases when local sensitivity makes patch application painful, the treatment patch should be applied on the forehead on the non painful side.

Turn the PainShield ON.



Notes

- Ensure that actuator (flat metal part) is in full contact with the skin.
- The PainShield may be used to improve penetration of topical ointment or gel.
- Each treatment actuator is effective for 30 full treatment sessions.